

32467



Trinity Biotech

Certificate of Analysis

Lot Number: 6956

Expiration Date: 2018-12-31

Production Date: 2016-12-09

Product Name: Premier Buffer A Reagent

REF: 01-03-0095 (940mL), 01-03-0080 (3.8L)

Shelf-Life: 2-Years (*Unopened with seal intact*), 30-Days (*After opening, if installed and capped.*)

Manufactured and Distributed By: Trinity Biotech, Kansas City, Missouri, USA

Storage Reqs: Keep tightly closed and store in a cool, dry location at 2-28°C, 36-82°F.

Use: For use with the Trinity Biotech Premier Hb9210 HbA1c Analyzer.

No substitutions or other uses are permitted.

Method of Analysis:

Visual Examination

Standard: A clear, colorless solution free from observable particulate matter

Result: The solution is clear, colorless and free from observable particulate matter.
Meets specification.

pH Conformance

Standard: 9.27 – 9.43

Result: Meets specification.

Control Performance

Standard: No test value outside of controls upper and lower limit on HPLC.

Result: Meets specification.

Approval

Quality Manager:

Date:

12/12/2016