

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: RUBELLA IgM - CHORUS

Code 81031

Lot No. 942 Expiry 2014-03

REAGENTS

Label	Description	Lot.	Exp	No
D2241	RUBELLA IgM	942	2014-03	6X6
B2241	CALIBRATORE RUBELLA IgM	942	2014-03	1
C2241	CONTROL SERUM	942	2014-03	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (INDEX)	1,6	1,1-2,0
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

1	3	5	9	4	2	5	0	0	0	0	0	6	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Be



CONTROL SERUM

1	3	5	9	4	2	5	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Be



The test has been carried out in accordance with the package insert

TESTED BY: Laura Capitani

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

Y. Bental

DATE 22-07-2013

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.