

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	GLUC-PAP
CATALOGUE NO:	GL8318
BATCH NO:	428143
EXPIRY DATE:	2019-10

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II 892UN	6.45	5.48-7.42	6.11	mmol/L
LEVEL III 766UE	16.1	13.7-18.5	15.8	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com