

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-11
KIT DESCRIPTION:	Uric Acid
CATALOGUE NO:	UA1613
BATCH NO:	397227
EXPIRY DATE:	2018-05

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II / 892UN	330	290-380	336	μmol/L
Level III / 766UE	550	480-630	556	μmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE