

Fisher Diagnostics a division of Fisher Scientific Company, LLC a part of Thermo Fisher Scientific Inc. 8365 Valley Pike Middletown, VA (USA) 22645-1905 Tel: 540 869 3200 or 800-528-0494

**Expiration Date** 

2

## Certificate of Analysis

**Lot Number** 

Product Name: Fibrinogen Assay Set - Orig:USA Expiration Date: 06/30/2019 Product Number: 100600 Storage: 2 - 8°C

Lot Number: 963358

**Description** 

Bovine Thrombin [BT] (2)	963355	2019-12-31
Imidazole Buffered Saline [IBS]	963357	2019-09-30
Fibrinogen Reference Plasma [FRP] (3)	963356	2019-06-30
Specification	Specification Limit	Result
BT Appearance Cake	White solid	Complies
BT Appearance Reconstituted	Clear, colorless with no particulates	Complies
BT pH	7.3 - 7.8	7.8
BT Moisture (Karl Fischer)	<= 3 %	0
BT Vial-to-Vial	<= 4 %	1
BT Assigned Value	100 NIH U/mL	100 NIH U/mL
IBS Appearance	Clear, colorless with no particulates	Complies
IBS pH	7.2 - 7.6	7.5
FRP Appearance Cake	Yellow to buff colored solid	Complies
FRP Appearance Reconstituted	Yellow to straw colored, clear to hazy,	Complies
	free of particulates when viewed against	
	dark background	
FRP Solubility	Less than 2 minutes	Complies
FRP Moisture (Karl Fischer)	<= 3 %	1
FRP Vial-to-Vial	<= 1.5 %	1.5
FRP Assigned Value	200 - 400 mg/dL	293
r-squared (photo-optical)	>= 0.96	0.99
Low Control	Assigned Value	107 mg/dL
Low Control	Recovered Value	110 mg/dL
Low Control (Diff from assigned)	-30 - 30 mg/dL	3
Normal Control	Assigned Value	292 mg/dL
Normal Control	Recovered Value	308 mg/dL
Normal Control (% of assigned)	-15 - 15 %	5
High Control	Assigned Value	635 mg/dL
High Control	Recovered Value	644 mg/dL
		_

For in-vitro diagnostic use.

High Control (% of assigned)

\* Clauss Fibrinogen

-15 - 15 %

PRINTED ON: 2017-01-10

<sup>1.</sup> All clot-based testing was performed on a photo-optical instrument.

<sup>2.</sup> The bovine based material used in this product is from farms certified to be free of transmissible spongiform encephalitis (TSE).



3. Each unit of source material used in the preparation of this product has been tested by an FDA licensed method and found non-reactive for HBsAg and negative for antibodies to HIV and HCV. However, no known test method can offer complete assurance that products derived from human blood will not transmit hepatitis, AIDS, or other infectious diseases. This product, like all materials of human origin, should be handled as potentially infectious biological material.

Quality Representative

Date: 2017-01-10

PRINTED ON: 2017-01-10