

DIESSE DIAGNOSTICA SENESE SpA

53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione

20144 Milano (Italy) Via A. Solari 19 sc. 6

Tel. +39-02-4859121 r.a. Fax +39-02-48008530

CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE**PRODOTTO/PRODUCT**

ESR CONTROL CUBE

REF:

10435 (4x9 ml)

10436 (2x9 ml)

REAGENTI/REAGENTS

	Kit	Level 1	Level 2
Lot:	251	251	251
	2016-12	2016-12	2016-12
	2017-12	2017-12	2017-12

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	3-10	80	55-105	OFF
VES-Matic Cube 80/VES-Matic Cube 200	7	3-10	50	25-75	OFF

Data/Date

20/03/2017

Testato da/Tested by:

Katia Tanzini

Approvato da/Approved by:

Grazia Dal Maso

(Responsabile CQ/QC Head)