

CERTIFICATE OF CONFORMITY

Certificate for Lot Number: 45317

Product Code: TS/5-40L Product: 5ml Letheen Brot Blue Bkpt/Visc Swab (Pack Size 100)

Date Of Manufacture: 09/2014 Expiry: 2016-09 Shelf Life: 24 months

This is to certify that the product detailed on this certificate conforms to Technical Service Consultants Ltd procedures and specifications and has performed the tests detailed below within documented parameters.

Sterilisation Reference, Isotron PLC Number

S11240764-1-1 - 1613703 : Dose 25.3 - 40.1 kGy

For Irradiated products batch release is controlled by the verification of the radiation dose throughout the load.

Autoclave Reference, Time Manufactured

For autoclaved products, either a temperature datalogger or visual sterilisation indicator are used to confirm that sterility was achieved.

Physical and Microbiological Assurance

Tests Performed	Results Obtained
Recovery of L.mono (160)	Pass
Sterility	Pass
Product Specification	Pass
Volume (4.9-5.2 ml)	Pass

Overall Result: Pass

Certificate Issued by: Joanne Walsh Certificate authorised on 29.09.2014

Electronically generated on 8th October 2014 - valid without signature

Sent to Vickie (vickie.bruckshaw@tscswabs.co.uk)