

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-11
KIT DESCRIPTION:	HbA1c
CATALOGUE NO:	HA3830
BATCH NO:	403268
EXPIRY DATE:	2018-05-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 1722HA	5.80	4.64-6.96	6.17	% HbA1c
LEVEL II / 1724HA	12.4	9.92-14.90	11.98	% HbA1c

APPROVED BY:

Lisa Reer

QUALITY CONTROL REPRESENTATIVE