

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: September 2017
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY 1 - LEVEL 3
CATALOGUE NO: IAS3115
LOT NUMBERS: 1656EC
EXPIRY DATE: 2018-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Anti TPO (IU/ml)	1656EC	122-204	163	190	N/A
				167	
Anti TG (IU/ml)	1656EC	291-485	388	379	
				391	
C-Peptide (nmol/l)	1656EC	1.910-3.180	2.540	2.747	
				2.618	

QC TESTED BY: Davina Graham
DATE: 08 Sep 2017
RESULT (PASS/FAIL): pass
QC APPROVED BY: Adele Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016