

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME: CLEANING SOLUTION 2000

Code	83609			
Lot No.	070	Expiry	2019-10	

REAGENTS

Label	Description	Lot.	Exp	No
B2570	CLEANING SOLUTION 2000	070	2019-10	4

CONTROLS

	Found results	Expected result
Density	1.052	1.040 - 1.060

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 17/05/2018

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

na

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.