

Certificate Of Analysis



Orphée S.A. assures that this reagent batch was manufactured under control of proper Quality Management System (ISO 13485).

Each batch undergoes a series of stringent validation and control tests

Product: Only One
Reference: HM22-002-1
Batch number: 1717

Use Before: 24-April-2019

Tested on:	April 25, 2017		
Tested by:	J. GIRAUD		
Tested with:	Mythic 22 Nb: 201908-000316		
Other Reagents:	ORPHEE HM22-001-1 Lot number:	1627700004	Exp date: 2019-11
	ORPHEE HM22-003-1 Lot number:	1624400009	Exp date: 2018-09
		L MYT1704	Exp date: 2017-06-05
		N MYT1704	Exp date: 2017-06-05
		H MYT1704	Exp date: 2017-06-05
		Specifications	Conclusion
		*Osmolarity:	335±10 Pass
	Physico-Chemical properties	*pH at 25°C:	6.55±0.1 Pass
		*Conductivity (mS.cm ⁻¹):	17.0±1.0 Pass
	Bio-Chemical efficiency	Performance Tests:	Pass

Conclusion:

Actual-value meets specifications

Date: 10.10.2017

Prepared by Anna Spryszak / QA

Date: 10.10.2017

Approved by Robert Wiśniewski / QMS Representative

PRINTED VERSION IS NOT SUPERVISED, CHECK TOPICALITY OF DOCUMENT BEFORE USE

Document number	Document version	Date of revision	Prepared by	Approved by
FOR036	2	27.10.2015	RW	JP