



Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

Specifications Microorganism Name: Pseudomonas aeruginosa Catalog Number: 0484 Lot Number: 484-950** Reference Number: ATCC® 9027™* Purity: Pure Passage from Reference: 2	Expiration Date: 2019/12/31 Release Information: Quality Control Technologist: Tracy A Blenker Release Date: 2018/1/30
---	---

Performance	
Macroscopic Features: Large, flat, circular to irregular shaped, gray with silver sheen. A second colony type may also be present as small, round, shiny colonies.	Medium: SBAP
Microscopic Features: Straight or slightly curved gram negative rod.	Method: Gram Stain (1)
ID System: MALDI-TOF (1) See attached ID System results document.	Other Features/ Challenges: Results (1) Oxidase (Kovacs): positive (1) Motility B Medium: positive Pseudomonas P Agar: positive (blue-green color diffusing into the agar) Growth at 42 C: positive  Amanda Kuperus Quality Control Manager AUTHORIZED SIGNATURE

**Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

Note for Vitek®: Although the Vitek® panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods.

⚠ Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.



(*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.

(1) These tests are accredited to ISO/IEC 17025:2005.



Bruker Daltonik MALDI Biotyper Classification Results



Meaning of Score Values

Range	Interpretation	Symbols	Color
2.00 - 3.00	High-confidence identification	(+++)	green
1.70 - 1.99	Low-confidence identification	(+)	yellow
0.00 - 1.69	No Organism Identification Possible	(-)	red

Meaning of Consistency Categories (A - C)

Category	Interpretation
(A)	High consistency: The best match is a high-confidence identification. The second-best match is (1) a high-confidence identification in which the species is identical to the best match, (2) a low-confidence identification in which the species or genus is identical to the best match, or (3) a non-identification.
(B)	Low consistency: The requirements for high consistency are not met. The best match is a high- or low-confidence identification. The second-best match is (1) a high- or low-confidence identification in which the genus is identical to the best match or (2) a non-identification.
(C)	No consistency: The requirements for high or low consistency are not met.

Sample Name: Pseudomonas aeruginosa
 Sample Description: 0484
 Sample ID: 484-950
 Sample Creation Date/Time: 2018-01-24T12:02:24.002 JS/MB
 Applied MSP Library(ies): BDAL, Mycobacteria Library (bead method), Filamentous Fungi Library 1.0, Listeria

Sample Name	Sample ID	Organism (best match)	Score Value
F10 (+++)(A)	484-950	Pseudomonas aeruginosa	2.52

Comments:

N/A