Certificate Of Analysis



Orphée S.A. assures that this reagent batch was manufactured under control of proper Quality Management System (ISO 13485).

Each batch undergoes a series of stringent validation and control tests

Product: Only One Reference: HM22-002-1 Use Before: 29 -May-2019 Batch number: 1722

Tested on: May 30, 2017 Tested by: J. GIRAUD

Tested with: Mythic 22 Nb: 201908-000316

Other Reagents: ORPHEE HM22-001-1 Lot number: 1627700004 Exp date: 2019-11

> ORPHEE HM22-003-1 Lot number: 1624400009 Exp date: 2018-09

> > L MYT1704

Exp date: 2017-06-05 Hematology Controls Lot number: Exp date: 2017-06-05 N MYT1704

H MYT1704 Exp date: 2017-06-05

> Specifications Conclusion

Osmolarity: 335±10 Pass

Physico-Chemical properties •pH at 25°C: 6.55 ± 0.1 **Pass**

> *Conductivity (mS.cm⁻¹): 17.0±1.0 **Pass**

Bio-Chemical efficiency Performance Tests: Pass

Conclusion:

Actual-value meets specifications

Date: 21.12.2017

Prepared by Anna Spryszak / RA&QA

Date: 21.12.2017

Approved by Robert Wiśniewski / QMS Representative

PRINTED VERSION IS NOT SUPERVISED, CHECK TOPICALITY OF DOCUMENT BEFORE USE

Document number	Document version	Date of revision	Prepared by	Approved by
FOR036	2	27.10.2015	RW	JP