

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	FULL RANGE CRP (frCRP)
CATALOGUE NO:	CP3847
BATCH NO:	433011
EXPIRY DATE:	2019-05-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I 450LPC	23.2	18.6-27.8	21.4	mg/L
LEVEL II 451LPC	46.1	36.9-55.3	42.3	mg/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Server Danuelly