



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-02
KIT DESCRIPTION:	MICROALBUMIN (mALB)
CATALOGUE NO:	MA3828
BATCH NO:	444474
EXPIRY DATE:	2019-10-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I 922MA	32.3	25.8 – 38.7	34.2	MG/L
LEVEL II 923MA	159.2	127.4 – 191.1	160.1	MG/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

4706-QC Rev4
27th September 2013