

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017 – 11
KIT DESCRIPTION:	LDH P-L
CATALOGUE NO:	LD3818
BATCH NO:	430812
EXPIRY DATE:	2019 – 02

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 892UN	443	376 – 510	443	U/L
LEVEL III / 766UE	812	690 - 934	774	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE