

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

DATE OF MANUFACTURE: 2015-09
KIT DESCRIPTION: AMMONIUM
CATALOGUE NO: AM1054
BATCH NO: 398214
EXPIRY DATE: 2017-08

SAMPLE	TARGET	RANGE	RESULT
LEVEL II 482UN	116	95.3-137	136
LEVEL III 355UE	351	288-414	398
LINEARITY:	1760 µmol/L		

APPROVED BY:

Andree Russell.

QUALITY CONTROL ANALYST