

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: VARICELLA IgG - CHORUS

Code 81078			
Lot No.	009	Expiry	2015-02

REAGENTS

Label	Description	Lot.	Exp	No
D2471	VARICELLA IgG	009	2015-02	6x6
B2472	CALIBRATORE VARICELLA IgG	009	2015-02	1
C2471	CONTROL SERUM VARICELLA IgG	009	2015-02	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	330	≥ 100
CONTROL SERUM (Index)	2,8	1,9-3,5
NEGATIVE CONTROL (Index)	0,0	<0,7

CALIBRATOR

0	1	4	0	0	9	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


0	1	4	0	0	9	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Laura Capitani

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 03-03-2014

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.