RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACT	URE:	April 2017		
KIT DESCRIPTION:	IA Special	lity II Level 1		
CATALOGUE NO:	IAS3117			
LOT NUMBERS:	1672EC			
EXPIRY DATE:	2018-10-28			
SYSTEM OF ANALYS	IS: Siemens	s Immulite 2000	XPI	

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pgm/l)	1672EC	16.2 - 27.0	21.6	18.3	N/A
				18.5	
Gastrin (pmol/l)	1672EC	16.9 - 28.1	22.5	21.4	N/A
				20.8	

QC TESTED BY:		James Breen		
DATE:	05 Apr 2017		_	
RESULT	(PASS/FAIL):	Pass	_	
QC APP	ROVED BY:	adèle Martin.	_	

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr. 5.P. Fitzgerald, BSc. Ph.D. DSc. (Hons). FIAE Man. R. J. Fitzgerald, W. Fitzgerald, OBE. R. P. Yelly. BSc. FCA. Randox: Laboratoriea Limited is a company registered in Northern Ireland, Company Number NII 15738, VAT. Number. GB. 253, 0304, 00.