

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-03		
KIT DESCRIPTION:	LACTATE DEHYDROGENASE P-L REAGENT		
CATALOGUE NO:	LD8322		
BATCH NO:	409054		
EXPIRY DATE:	2018-05		

SAMPLE	TARGET	RANGE	RESULT	UNITS
892UN / LEVEL II	443	376 – 510	448	U/L
766UE / LEVEL III	812	690 - 934	761	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com