



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-02
KIT DESCRIPTION:	DIRECT BILIRUBIN
CATALOGUE NO:	BR8308
BATCH NO:	396709
EXPIRY DATE:	2017-02

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL 2/ 499DC	5.37	4.30-6.44	5.72	µg/ml
LEVEL 3/ 502DC	8.33	6.67-9.99	8.87	µg/ml

APPROVED BY:

Neil McFarlane

QUALITY CONTROL ANALYST