

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: May 2018
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS TRI LEVEL
CATALOGUE NO: IA3112
LOT NUMBERS: 453660 (1752EC, 1764EC, 1766EC)
EXPIRY DATE: 2020-10-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
C-PEPTIDE (nmol/l)	1752EC	0.278 - 0.463	0.37	0.394	N/A
				0.397	
LUTEINISING HORMONE (U/l)	1752EC	4.0 - 6.0	5.0	5.13	N/A
				4.95	
PROLACTIN (mU/l)	1752EC	144 - 216	180	186	N/A
				169	
C-PEPTIDE (nmol/l)	1764EC	1.32 - 2.20	1.76	1.801	N/A
				1.811	
LUTEINISING HORMONE (U/l)	1764EC	30.6 - 46.0	38.3	35.7	N/A
				36.5	
PROLACTIN (mU/l)	1764EC	661 - 991	826	848	N/A
				835	
C-PEPTIDE (nmol/l)	1766EC	2.29 - 3.81	3.05	2.860	N/A
				3.148	
LUTEINISING HORMONE (U/l)	1766EC	48.7 - 73.1	60.9	58.9	N/A
				62.4	
PROLACTIN (mU/l)	1766EC	1283 - 1925	1604	1605	N/A
				1618	

QC TESTED BY: Nicola McNamara
DATE: 09-May-18
RESULT (PASS/FAIL): Pass

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number: GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016