

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-02
KIT DESCRIPTION:	DIRECT BILIRUBIN
CATALOGUE NO:	BR8035
BATCH NO:	404544
EXPIRY DATE:	2019-01

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 1126UN	18.9	14.9-22.9	17.1	μmol/L
LEVEL III / 766UE	30.0	23.7-36.3	30.8	μmol/L

APPROVED BY:

Lisa Reu.

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited. 55 Diamond Road. Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com