

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-02
KIT DESCRIPTION:	UREA
CATALOGUE NO:	UR457
BATCH NO:	406699
EXPIRY DATE:	2020-01

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 892UN	7.05	5.99-8.11	7.05	mmol/L
LEVEL III/ 766UE	19.2	16.3-22.1	18.6	mmol/L

APPROVED BY:

h. Pandiorani

QUALITY CONTROL REPRESENTATIVE