

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: April 2018  
KIT DESCRIPTION: IA SPECIALITY 1 LEV 1  
CATALOGUE NO: IAS3113  
LOT NUMBERS: 1772EC  
EXPIRY DATE: 2019-12-28  
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

### TEST CONTROLS OR VIALS

| ANALYTE         | LOT NUMBER | RANGE       | TARGET | RESULT | PRECISION (if applicable) |
|-----------------|------------|-------------|--------|--------|---------------------------|
| Anti TG (kU/l)  | 1772EC     | 75 - 125    | 100    | 90.2   | N/A                       |
|                 |            |             |        | 95.1   |                           |
|                 |            |             |        |        |                           |
| Anti TPO (kU/l) | 1772EC     | 26.1 - 43.5 | 34.8   | 34.1   | N/A                       |
|                 |            |             |        | 33.3   |                           |
|                 |            |             |        |        |                           |
|                 |            |             |        |        |                           |
|                 |            |             |        |        |                           |
|                 |            |             |        |        |                           |

QC TESTED BY: Nicola McNamara  
DATE: 27-Apr-18  
RESULT (PASS/FAIL): Pass  
QC APPROVED BY: \_\_\_\_\_

*Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Man. R J Fitzgerald, W Fitzgerald OBE R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016