



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	BLOOD GAS CONTROL
CATALOGUE NO:	BG5003
BATCH NO:	177BG
EXPIRY DATE:	2019-02

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL III 177BG	SODIUM	164	156 – 172	171	mmol/L
LEVEL III 177BG	CHLORIDE	126	116 – 136	120	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hon) FAE MSc R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI5738, VAT Number GB 353 0304 00.