



Trinity Biotech

Certificate of Analysis

Lot Number: 7439

Expiration Date: 2019-05-31

Production Date: 2017-05-15

Product Name: Premier Buffer B Reagent

REF: 01-03-0096 (940ml), 01-03-0081 (3.8L)

Shelf-Life: 2-Years (*Unopened with seal intact*), 30-Days (*After opening, if installed and capped.*)

Manufactured and Distributed By: Trinity Biotech, Kansas City, Missouri, USA

Storage Reqs: Keep tightly closed and store in a cool, dry location at 2-28°C, 36-82°F.

Use: For use with the Trinity Biotech Premier Hb9210 HbA1c Analyzer.
No substitutions or other uses are permitted.

Method of Analysis:

Visual Examination

Standard: A clear, colorless solution free from observable particulate matter

Result: The solution is clear, colorless and free from observable particulate matter.
Meets specification.

Baseline Flatness

Standard: Flat baseline with no deflection.

Result: The baseline is flat with no deflection. Meets specification.

Control Performance

Standard: No test value outside of controls upper and lower limit on HPLC.

Result: Meets specification.

pH Conformance

Standard: 8.00 - 9.00

Result: Meets specification.

Approval

Quality Manager:

Date:

5/19/2017