



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-07
KIT DESCRIPTION:	LDH P-L
CATALOGUE NO:	LD3818
BATCH NO:	419506
EXPIRY DATE:	2018-09

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 892UN	443	376-510	451	U/L
LEVEL III/ 766UE	812	690-934	770	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald CBE, R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5738, VAT Number GB 353 0304 00

4706-QC Rev4
27th September 2013