

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-11
KIT DESCRIPTION:	CREATININE REAGENT
CATALOGUE NO:	CR8316
BATCH NO:	396783
EXPIRY DATE:	2019-10

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 892UN	121	97 - 145	125	μmol/L
LEVEL III / 766UE	320	256 - 384	315	μmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com