

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: Epstein-Barr VCA IgG- CHORUS

Code 81055		
Lot No. 353	2018-10	2017-07

REAGENTS

Label	Description	Lot.	Exp	No
D2341	Epstein-Barr VCA IgG	353	2018-10	6X6
B2342	CALIBRATORE Epstein-Barr VCA IgG	353	2018-10	1
C2341	CONTROL SERUM Epstein-Barr VCA IgG	353	2018-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	305	≥ 100
Control Serum (INDEX)	3,9	2,8-5,2
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

0	0	5	3	5	3	7	0	0	0	0	0	2	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Alianali*



CONTROL SERUM

0	0	5	3	5	3	7	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Alianali*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Alianali*

DATE 11-08-2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.