

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

DATE OF MANUFACTURE: 2015-08
KIT DESCRIPTION: AST
CATALOGUE NO: AS1204
BATCH NO: 390246
EXPIRY DATE: 2017-07

SAMPLE	TARGET	RANGE	RESULT
LEVEL II 482UN	116	95.3-137	136
LEVEL III 355UE	351	288-414	398
LINEARITY:	1760 µmol/L		

APPROVED BY:

Andree Russell.

QUALITY CONTROL ANALYST