



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-07
KIT DESCRIPTION:	URIC ACID
CATALOGUE NO:	UA230
4BATCH NO:	420886
EXPIRY DATE:	2020-06

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II 892UN	330	290-380	316	μmol/L
LEVEL III 766UE	550	480-630	534	μmol/L

APPROVED BY:

Louise McStravick

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R F Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5758, VAT Number GB 353 0304 00

4706-QC Rev4
27th September 2013