

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

DATE OF MANUFACTURE: 2015-11
KIT DESCRIPTION: ALT
CATALOGUE NO: AL8304
BATCH NO: 403090
EXPIRY DATE: 2016-12

SAMPLE	TARGET	RANGE	RESULT
LEVELII 482UN	0.32	0.27-0.37	0.290mmol/l
LEVELIII 355UE	1.51	1.28-1.74	1.35mmol/l
LINEARITY:	3.2mmol/l		

APPROVED BY:

U. Pandiarani

QUALITY CONTROL ANALYST