

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	LIQUID PROTEIN CONTROL 2
CATALOGUE NO:	PS2683
BATCH NO:	496LPC
EXPIRY DATE:	2020-05

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
496LPC	AAG	77.9	62.3-93.5	73.5	mg/dL
496LPC	AAT	143	114-172	140	mg/dL

**APPROVED BY:** 

h. Pandiciani

**QUALITY CONTROL REPRESENTATIVE**