



CERTIFICATE OF BATCH ANALYSIS REPORT

Diesse Diagnostica Senese S.p.A hereby Certifies

that the product: **VACU-TEC** Code: **10200/W**

Lot. N.: **7324** Exp. Date: **2018-09**

Prod. Date: **2017-11**

Packed in the pallet N.: **5042**

have been approved after testing the following controls.

ON LINE QUALITY CONTROLS (model CDP not enclosed)

- sodium citrate solution concentration
- sodium citrate volume
- vacuum level (*)
- correct placement of the cap
- no extra body inside the product

(*) Only for VACU-CODE and VACU-TEC tubes

QC Passed



PACKAGING QUALITY CONTROL (model CDP not enclosed)

- control of the number of the product for each pack

Pack N.: **100**

QC Passed



QUALITY CONTROLS ON FINISHED PRODUCT (model CDP not enclosed)

- pack integrity
- data printed on the product label

QC Passed



Production Manager Signature

Date: 2017-11-22

Document Number:

QC Manager Signature

Date: 2017-11-22

ODV **72084**

Stamp and signature