

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-10
KIT DESCRIPTION:	DIRECT HDL CHOLESTEROL (HDL)
CATALOGUE NO:	CH8311
BATCH NO:	468323
EXPIRY DATE:	2020-01-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I 2530CH	0.72	0.61 – 0.82	0.77	mmol/l
LEVEL II 2532CH	1.25	1.06 – 1.44	1.35	mmol/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE