

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-12
KIT DESCRIPTION:	C3
CATALOGUE NO:	CM3845
BATCH NO:	403698
EXPIRY DATE:	2018-09-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LPC1 453LPC	74.4	59.5 – 89.3	70.9	mg/dl
LPC3 455LPC	227.0	182.0 – 272.0	221.7	mg/dl

APPROVED BY:

Saah M'Coy.

QUALITY CONTROL REPRESENTATIVE