



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Radox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |            |
|----------------------|------------|
| DATE OF MANUFACTURE: | 2017-02    |
| KIT DESCRIPTION:     | ALT        |
| CATALOGUE NO:        | AL3801     |
| BATCH NO:            | 406578     |
| EXPIRY DATE:         | 2018-05-28 |

| SAMPLE               | TARGET | RANGE   | RESULT | UNITS |
|----------------------|--------|---------|--------|-------|
| Level II /<br>892UN  | 38     | 30-46   | 35     | U/I   |
| Level III /<br>766UE | 144    | 115-173 | 140    | U/I   |

**APPROVED BY:**

David Richmond

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIACI, Mr F J Fitzgerald, W Fitzgerald CBE, R F Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5739, VAT Number GB 353 0304 00.

4706-QC Rev4  
27<sup>th</sup> September 2013