



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |                          |
|----------------------|--------------------------|
| DATE OF MANUFACTURE: | 2017-10                  |
| KIT DESCRIPTION:     | LIQUID PROTEIN CONTROL 3 |
| CATALOGUE NO:        | PS2684                   |
| BATCH NO:            | 498LPC                   |
| EXPIRY DATE:         | 2020-05                  |

| SAMPLE | ANALYTE       | TARGET | RANGE     | RESULT | UNITS |
|--------|---------------|--------|-----------|--------|-------|
| 498LPC | COMPLEMENT C3 | 251    | 201-301   | 241    | mg/dL |
| 498LPC | COMPLEMENT C4 | 51.1   | 40.9-61.3 | 49.3   | mg/dL |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.