

CERTIFICATE OF ANALYSIS

Reagent: Creatinine-S
Catalogue Number: 221-30, 221-50
Lot Number: 51320
Expiry Date: 2020-04-30

A. PHYSICAL CHARACTERISTICS

PARAMETER	SPECIFICATION	RESULT
Optical Density of Picric Acid Reagent at 510 nm at 25°C:	0.053-0.057	0.055
Linearity	20 mg/dL	Pass
Performance	Values obtained for normal and abnormal level control sera are within the ranges established for the method.	

B. CHEMICAL CHARACTERISTICS

1. Creatinine Base Reagent (R1): a solution containing 0.25 mol/L sodium hydroxide and surfactants.
2. Creatinine Picrate Reagent (R2): a solution containing 31.4 mmol/L picric acid.
3. Creatinine Standard: a solution containing 354 µmol/L (4.0 mg/dL) creatinine and a preservative (not included with Cat. No. 221-50). (For use with manual assay only).

The above reagent has passed all of our quality control procedures and has been released for distribution.

24-Jul-2017
Date

S. M. Jallo
Quality Control

CAT-COA-0108-3.0-Sekisui-Creatinine221

Sekisui Diagnostics, LLC

4 Hartwell Place
Lexington, MA 02421
Tel: 800-332-1042 Fax: 800-762-6311

Sekisui Diagnostics (UK) Limited

Liphook Way
Allington, Maidstone
Kent, ME16 0LQ UK
Tel: +44 (0)1622 607800 Fax: +44 (0)1622 607801

Sekisui Diagnostics P.E.I. Inc.

70 Watts Avenue
Charlottetown, Prince Edward Island
C1E 2B9 Canada
Tel: 800-565-0265 Fax: 902-628-6504