


RANDEX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

DATE OF MANUFACTURE: 2016-01
KIT DESCRIPTION: CALCIUM
CATALOGUE NO: CA3871
BATCH NO: 389800
EXPIRY DATE: 2017-07

SAMPLE	TARGET	RANGE	RESULT
LEVEL II 482UN	116	95.3-137	136
LEVEL III 355UE	351	288-414	398
LINEARITY:	1760 µmol/L		

APPROVED BY:



QUALITY CONTROL ANALYST