

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04		
KIT DESCRIPTION:	DIRECT BILIRUBIN REAGENT		
CATALOGUE NO:	BR8308		
BATCH NO:	410645		
EXPIRY DATE:	2018-04		

SAMPLE	TARGET	RANGE	RESULT	UNITS
1127UN / LEVEL II	15.9	12.6 – 19.2	14.8	μmol/L
766UE / LEVEL III	31.6	25.0 – 38.2	31.2	μmol/L

**APPROVED BY:** 

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com