

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	21st January 2016
KIT DESCRIPTION:	Alkaline Phosphatase
CATALOGUE NO:	AP3802
BATCH NO:	397104
EXPIRY DATE:	2017-01

SAMPLE	TARGET	RANGE	RESULT	UNITS
Therapeutic drug control Level II (499DC)	9.01	7.21 – 10.80	8.40	µg/ml
Therapeutic drug control Level III (524DC)	18.70	15.00 – 22.40	17.65	µg/ml

APPROVED BY:

Nathan Stewart

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number: GB 353 0304 00.

