



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2015 – 11
KIT DESCRIPTION:	ALKALINE PHOSPHATASE
CATALOGUE NO:	AP3803
BATCH NO:	400348
EXPIRY DATE:	2016 - 12

SAMPLE	TARGET	RANGE	RESULT	UNITS
DRUG CONTROL LEVEL II / 499DC	1.73	1.38 – 2.08	1.93	ng/ml
DRUG CONTROL LEVEL III / 502DC	3.34	2.67 – 4.01	3.15	ng/ml

APPROVED BY:

QUALITY CONTROL ANALYST