

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2012 - 10
KIT DESCRIPTION:	LIQUID CARDIAC CONTROL LEVEL 2
CATALOGUE NO:	CQ5052
BATCH NO:	3459CK
EXPIRY DATE:	2014 - 10

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 3459CK	HSCRP	1.97	1.58 – 2.36	2.04	mg/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE





RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2012 - 10
KIT DESCRIPTION:	LIQUID CARDIAC CONTROL LEVEL 3
CATALOGUE NO:	CQ5053
BATCH NO:	3457CK
EXPIRY DATE:	2014 - 10

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL III / 3457CK	HSCRP	8.09	6.47 – 9.71	8.54	mg/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com



13th March 2013