

Certificate of Analysis

Product Name: Reference Emulsion Fluid - Orig:USA
Product Number: 100240
Lot Number: 485760

Expiration Date: 10/31/2015
Storage: Room Temp

Specification	Specification Limit	Result
Coagulation Optics Test (ACL Index)	-8 - 8 UNITS	6
Coagulation Control Level 1 PT	-0.9 - 0.9 S	0.1
Coagulation Control Level 3 PT	-6.0 - 6.0 S	-0.9
Coagulation Control Level 1 APTT	-3.0 - 3.0 S	-0.2
Coagulation Control Level 3 APTT	-9.0 - 9.0 S	-1.5
Normal Fibrinogen	-50 - 50 mg/dL	-2
Low Fibrinogen	-60 - 60 mg/dL	9

For in-vitro diagnostic use.

Storage Restriction: Do Not Freeze.

All results displayed above are expressed in comparison to the Master.



Quality Representative

Date: 2014-05-01