

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFAC	TURE:	October 2017	
KIT DESCRIPTION:	IA Special	ity II level 2	
CATALOGUE NO:	IAS3118		
LOT NUMBERS:	1701EC		
EXPIRY DATE:	2019-03-28		
SYSTEM OF ANALY	SIS: Siemens	Immulite 2000	

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
	1701EC	95.1 - 159	127	123	N/A
Calcitonin (pg/ml)				122	
Gastrin (pmol/l)	1701EC	61 - 102	81.3	95.3	N/A
				85.3	
		4.00			

QC TEST	ED BY:	Ilona Palmer		
DATE:	20 Oct 2017			
RESULT	(PASS/FAIL):	Pass		
QC APPR	ROVED BY:	adèle !	Martin.	