

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2015-08
KIT DESCRIPTION:	HIGH CONTROL BILIRUBIN
CATALOGUE NO:	BE454
BATCH NO:	066BE
EXPIRY DATE:	2018-01

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I 196LPC	1.74	1.39-2.09	1.68	mg/L
LEVEL II 197LPC	3.73	2.98-4.48	3.41	mg/L

APPROVED BY: *Louise McStravick*

QUALITY CONTROL REPRESENTATIVE