

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-03
KIT DESCRIPTION:	DIRECT HDL-CHOLESTEROL
CATALOGUE NO:	CH3811
BATCH NO:	445989
EXPIRY DATE:	2019-07-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 2532CH	1.25	1.06-1.44	1.28	mmol/L
LEVEL III / 2534CH	1.53	1.30-1.76	1.52	mmol/L

APPROVED BY:

vette Milla

QUALITY CONTROL REPRESENTATIVE