

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	UREA
CATALOGUE NO:	UR3825
BATCH NO:	426381
EXPIRY DATE:	2018-12

SAMPLE	TARGET	RANGE	RESULT	UNITS
HUMAN LEVEL II 892UN	7.04	5.98-8.10	7.05	mmol/L
HUMAN LEVEL III 766UE	19.2	16.3-22.1	18.9	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE